

### Amendments to Claims

#### Please amend the claims as follows:

In agreement with Examiner's suggestion to remove an antecedent basis error in

5 Claim 15, please exchange the article 'the' for the article 'a' as indicated herefollowing.

15) Optical in-vivo monitoring systems of claim 11, further comprising:

an opaque region;

a lens;

10 a computer; and

indicator in [the] a watch face,

said opaque region being disposed on a central portion of the thin-disk

bottom surface whereby light from the illumination source is blocked from

entering the thin-disk unless it has passed first through a sufficient amount of

15 tissue,

said lens being coupled with a semiconductor light source to form an  
optical illumination beam symmetric about a system axis whereby light from the  
source is transmitted into tissue at a test site,

20 said computer is coupled to said photodetector whereby electric signals  
produced at the photodetector can be analyzed and processed in accordance with  
appropriate software running thereon and the computer is further coupled to the  
indicator whereby information may be graphically presented to a user.

In agreement with Examiner's suggestion regarding quotation marks of claim 19,

25 please omit these marks as follows:

19) Methods of in-vivo optical monitoring of claim 17, said ["]concentrating  
beam["] step is further defined as causing light which enters an exit aperture to be  
transmitted through a sufficiently smaller exit aperture coupled to a photodetector.

PO Box 757, La Jolla, CA 92038  
858 699 6015; [jpage@integrityip.com](mailto:jpage@integrityip.com)

**Claims pending after current amendment is applied**

- 5            1)        (Original)     A thin-disk optical coupling element comprising:  
                 an entrance aperture;  
                 a beam turning element;  
                 a waveguide array; and  
                 an exit aperture,  
10            said entrance aperture comprising a planar annular region of a bottom surface of  
                 said thin-disk,  
                 said beam turning element formed in or on said thin-disk is arranged to redirect  
                 beams incident substantially orthogonal thereon in a radially inward direction towards  
                 said waveguide array,  
15            said waveguide array is formed in a central or inner portion of said thin-disk, the  
                 waveguide array comprising a plurality of pie-wedge shaped members, and  
                 said exit aperture being at least one surface prepared such that light beams from  
                 said waveguide array passes there through and exits the thin disk at an appreciably higher  
                 energy density than the beams energy density at entry to the thin-disk.  
20            2)        (Original)     A thin-disk optical coupling of claim 1, said beam turning  
                 element is further characterized as a grating of periodic structures arranged with circular  
                 symmetry, said structures being disposed on a thin-disk surface opposite the surface in  
                 which the entrance aperture lies.  
25            3)        (Original)     A thin-disk optical coupling of claim 2, said grating being a  
                 cylindrically symmetric blazed type grating formed of a plurality of surface relief repeat  
                 structures in a top surface of said thin-disk, each repeat structure forming a concentric  
                 circle having a radius different each of the other repeat structures.

PO Box 757, La Jolla, CA 92038  
858 699 6015; [jpage@integrityip.com](mailto:jpage@integrityip.com)

In re Application of: Florian, J.  
Serial Number: 10/780,813

4) (Original) A thin-disk optical coupling of claim 2, said grating being a symmetric phase type grating characterized as a hologram optical element.

5) (Original) A thin-disk optical coupling of claim 1, each of said entrance and exit apertures, said beam turning element, and said waveguide array being symmetric about a system symmetry axis,

said entrance aperture is polished smooth, the aperture being about between .5 and 5 cm<sup>2</sup>, having an opaque void at its center, and

said exit aperture is comprised of a plurality of curved surfaces lying in a cylinder.

6) (Original) A thin-disk optical coupling of claim 1, said waveguide array comprised of pie-wedge elements is further defined as having side surfaces and air slits between those side surfaces to provide interfaces operable for promoting total internal reflections within the pie-wedge elements, said side surfaces being planar polished surfaces whereby light beams falling thereon remain within the pie-wedge elements and continue to propagate generally towards the system symmetry axis.

7) (Original) A thin-disk optical coupling of claim 1, further comprising a mirror element; said mirror element being coupled with said waveguide array whereby light from the waveguide array falls upon the mirror element and is reflected toward said exit aperture.

8) (Original) A thin-disk optical coupling of claim 7, said mirror is a conic surface formed directly on the thin-disk element and concentric with the system axis.

9) (Original) A thin-disk optical coupling of claim 8, said exit aperture is formed on the top surface of the thin-disk element, the opposite surface with respect to

PO Box 757, La Jolla, CA 92038  
858 699 6015; [jpage@integrityip.com](mailto:jpage@integrityip.com)

In re Application of: Florian, J.  
Serial Number: 10/780,813

the entrance aperture, whereby light leaves the thin disk on the axis and upwardly away from the tissue test site.

10) (Original) A thin-disk optical coupling of claim 1, said coupling is  
5 operable for condensing light in accordance with an entrance to exit aperture ratio greater than 3.

11) (Original) Optical in-vivo monitoring systems for monitoring states of  
living tissues comprising:

10 an illumination source;  
an optical coupling element; and  
a photodetector,

said illumination source arranged to transmit a beam into a tissue test site, said  
optical coupling element concentric therewith said illumination source, being arranged to  
15 receive modulated light from the tissue test site and transmit received light at a  
sufficiently higher energy density to said photodetector.

12) (Original) Optical in-vivo monitoring systems of claim 11, said  
optical coupling element is further defined as an optical collection and condensing  
20 element having an entrance aperture and an exit aperture whose area ratio exceeds about  
3.

13) (Original) Optical in-vivo monitoring systems of claim 11, said  
photodetector is arranged on axis with said illumination source, the illumination source  
25 lying between the tissue test site and the photodetector.

14) (Original) Optical in-vivo monitoring systems of claim 11, said  
optical coupling element is further defined as a thin-disk optical coupling element  
comprising: an entrance aperture; a beam turning element; a waveguide array; and an exit

PO Box 757, La Jolla, CA 92038  
858 699 6015; [jpage@integrityip.com](mailto:jpage@integrityip.com)

In re Application of: Florian, J.  
Serial Number: 10/780,813

aperture, said entrance aperture comprising a planar annular region of a bottom surface of said thin-disk, said beam turning element formed in or on said thin-disk is arranged to redirect beams incident substantially orthogonal thereon in a radially inward direction towards said waveguide array, said waveguide array is formed in a central or inner portion of said thin-disk, the waveguide array comprising a plurality of pie-wedge shaped members, and said exit aperture being at least one surface prepared such that light beams from said waveguide array passes there through and exits the thin disk at an appreciably higher energy density than the beams energy density at entry to the thin-disk.

15) (Currently Amended) Optical in-vivo monitoring systems of claim 11, further comprising:

an opaque region;

a lens;

a computer; and

indicator in a watch face,

said opaque region being disposed on a central portion of the thin-disk bottom surface whereby light from the illumination source is blocked from entering the thin-disk unless it has passed first through a sufficient amount of tissue,

said lens being coupled with a semiconductor light source to form an optical illumination beam symmetric about a system axis whereby light from the source is transmitted into tissue at a test site,

said computer is coupled to said photodetector whereby electric signals produced at the photodetector can be analyzed and processed in accordance with appropriate software running thereon and the computer is further coupled to the indicator whereby information may be graphically presented to a user.

In re Application of: Florian, J.  
Serial Number: 10/780,813

16) (Original) Optical in-vivo monitoring systems of claim 11, said lens is arranged to form an illumination beam into a cone whereby light enters a tissue test site and tends to scatter radially outwardly from the system symmetry axis.

5 17) (Original) Methods of in-vivo optical monitoring, comprising the steps:  
illuminating tissue test site via an emitting semiconductor;  
receiving light which has sufficiently interacted with tissue via large scatter path  
length at a large annular aperture;  
10 turning light received at said aperture radially inward toward a symmetry axis;  
concentrating beam via TIR reflections in pie-wedge light pipe array; and  
converting received modulated optical beam to electrical impulses at a  
photodetector.

15 18) (Original) Methods of in-vivo optical monitoring of claim 17, further comprising the step extinguishing light which has interacted with tissue for less than an optical path of 5 mm at an opaque beam block concentric with said illumination source.

19) (Currently Amended) Methods of in-vivo optical monitoring of claim 17,  
20 said concentrating beam step is further defined as causing light which enters an exit aperture to be transmitted through a sufficiently smaller exit aperture coupled to a photodetector.

20) (Original) Methods of in-vivo optical monitoring of claim 17, further  
25 comprising the step: processing said electrical impulses to determine the state of said tissue site including a pulse rate.